

02220

STAT

November 25, 1968

STAT

STAT

STAT

Subject: Amendment to Specification for
 [redacted] High Power Stereo
 Comparator Head dated October 1, 1968

Dear [redacted]

Amendment #1 in our letter dated November 22, 1968, is incorrectly worded. The word "maximum" was inadvertently used in the first sentence of Amendment #1 instead of minimum. This letter is a written confirmation of our telegraphed request that the word maximum be removed from Amendment #1 and replaced with the word minimum.

STAT

STAT

[redacted] is hereby directed to submit their written quotation to the [redacted] based on the corrected wording of Amendment #1 which when changed in accordance with our telegram of November 25, 1968, will read in its entirety as follows:

Amendment #1

The final paragraph of the specification will be amended to read:

"With the 10X [redacted] eyepieces, the zoom at 2X, and the 10X [redacted] Fluotar objectives, the instrument will have a minimum axial resolution of approximately 1000 lines per millimeter. Under the same conditions, with the 6X [redacted] Fluotars, approximately 630 lines per millimeter and with the 3X [redacted] Fluotar, approximately 320 lines per millimeter."

STAT

STAT

STAT

Thank you for your cooperation in this matter.

Very truly yours,

STAT

[redacted]
 Asst. Manager, Engineering

STAT